Dentsply Sirona

Explanation of symbols used by Dentsply Sirona

	Cavity width (a) must be less than 2/3 of the intercuspal distance (b)
	Agitate adhesive slightly with applicator tip
	5 sec Remove solvent
	Curing light energy output and curing time instruction
	Curing light energy output
	Curing light energy output
	Curing time
	Curing time

	Curing time [s]
	Curing time (s) and maximum layer thickness (mm) of filling material to allow proper curing
	Maximum layer thickness (mm) of filling material to allow proper curing
	Maximum layer thickness (mm) of filling material to allow proper curing
	Battery disposal
	Opened packages are not replaced
	Atmospheric pressure
	Humidity limitation

© Dentsply Sirona 2019-04-09

	Caution (IFU)		Tri-layer bottle with flip-top-cap
	Information (IFU)		Single Unit Dose
Content 0.28 g	Compules® Tips		Bottle
	Syringe		ITG Illustrated Technique Guide
	Compules® Tips Gun		IFU Instructions for Use
	Compula® Tip		i-shade label
	Compules® Tip	CLOUD	Cloud shades
	CliXdish™	1 x	Shade guide
	Applicator Tip Flocked Applicator Tip		Conditioner syringe

© Dentsply Sirona 2019-04-09

	Syringe		Adhesive bottle
	Syringe Tip		Water dispensing bottle
	Powder Glass	1 x	1 Drop
	Liquid-Glass	1 x	1 Scoop
	Powder measure		Dosage: scoops : drops
	Capsule		Mixing pad
	Capsule Extruder		Applicator Needles
	Automix Syringe		X-Post
	Intra Oral Tips		Paper Points

© Dentsply Sirona 2019-04-09

	Peeso Reamer		Precision Drill
	Rotation		Composite Shade Guide Holder
Dosage 	Level Powder without Pressure	Handling 	Loosen Powder in Tub turning the box over at least two times
	2 x 25 ml Powder		Mix max. 30 Sec.
	34 ml H_2O		max. 1 Min. working time
	Medium Tray		min. 1 Min. setting time
EXP	Expiry / Expiration Date	MD	Medical Device according to European Medical Device Regulation (MDR)

© Dentsply Sirona 2019-04-09