Dentsply Sirona

Explanation of symbols used by Dentsply Sirona

	Cavity width (a) must be less than 2/3 of the intercuspal distance (b)
	Agitate adhesive slightly with applicator tip
	5 sec Remove solvent
	Curing light energy output and curing time instruction
	Curing light energy output
	Curing light energy output
	Curing time
	Curing time

	Curing time [s]
	Curing time (s) and maximum layer thickness (mm) of filling material to allow proper curing
	Maximum layer thickness (mm) of filling material to allow proper curing
	Maximum layer thickness (mm) of filling material to allow proper curing
	Battery disposal
	Opened packages are not replaced
	Atmospheric pressure
	Humidity limitation

© Dentsply Sirona 2020-02-04

	Caution (IFU)		Syringe Tip
	Information (IFU)		Applicator Tip Flocked Applicator Tip
	CliXdish™		Applicator Needles
	Compules® Tips Gun		ITG Illustrated Technique Guide
	Compula® Tip		IFU Instructions for Use
	Compules® Tip		i-shade label
	Syringe	CLOUD	Cloud shades
	Conditioner syringe	1 x	Shade guide
	Syringe	2 x	Composite Shade Guide Holder

© Dentsply Sirona 2020-02-04

	Tri-layer bottle with flip-top-cap		Adhesive bottle
	Single Unit Dose		Water dispensing bottle
	Bottle	1 x	1 Drop
	Liquid-Glass	1 x	1 Scoop
	Powder Glass	:	Dosage: scoops : drops
	Capsule		Mixing pad
	Capsule Extruder		Powder measure
	Automix Syringe		X-Post
	Intra Oral Tips		Paper Points

© Dentsply Sirona 2020-02-04

	Peeso Reamer
	Rotation
Dosage 	Level Powder without Pressure
	2 x 25 ml Powder
	34 ml H_2O
	Medium Tray
EXP	Expiry / Expiration Date

	Precision Drill
Handling 	Loosen Powder in Tub turning the box over at least two times
	Mix max. 30 Sec.
	max. 1 Min. working time
	min. 1 Min. setting time
MD	Medical Device according to European Medical Device Regulation (MDR)

© Dentsply Sirona 2020-02-04